



# Medical gas devices



## RVB collection jar

### Product description



Organismo Notificato n. 0476  
Notified Body nr. 0476

The collection jar is installed upstream of the RVB vacuum regulator to protect the suction system and regulator from impurities; it has a hose connector on the inlet and is connected to the vacuum regulator by a G 1/2" F R thread; the collection jar is made of polycarbonate and has a capacity of 150 cc or 300 cc; the cap has an outlet with a hose connector; the RVB jar (collection jar, lid, O-ring) can be periodically placed in an autoclave up to a temperature of 134°C and each time it is used by a different patient for a maximum of 100 cycles at

121°/126 ° or 85 cycles at 134°, for a maximum total cycle time of 10 min. (of which 6 cycles are performed at 134°);

### Normatives

UNI EN ISO 10079-3 | UNI EN ISO 15001 | UNI EN ISO 10079-4

### Components

- Makrolon jar.
- Yellow PVC top.
- Moplen float.
- One hose connector inlet.
- N.One outlet with G 1/2" F R rotary nut.
- EPDM PEROX O-ring.

### Technical data

CODE	VOLUME	CONNECTION	WEIGHT
HR081	150 cc	G1/2" M R	0.78 kg
HR082	300 cc	G1/2" M R	0.84 kg

