



Medical gas devices

UNI 9507 PROBE



Organismo Notificato n. 0476
Notified Body nr. 0476

Product description



The UNI probe complies with UNI 9507 standards; the probe can be used to supply the desired medical gas from the clinic-hospital system. Each probe is a different size so they cannot be swapped and each individual probe can be assigned to the corresponding gas take-off. All materials are oxygen compatible, supplied clean and free of any traces of oil, grease or dust.

Normatives

UNI EN ISO 9170-1 | NF S 90-116 | UNI 9507 | UNI EN ISO 9170-2 | UNI EN ISO 15001 | FD S 90-119

Components

Body made of CW617N chemically nickel-plated brass.
Specific label for each gas.

Technical data

CODE	GAS	OUTLET	CONNECTION	WEIGHT
HI010	O ₂	In line	Hose connector Ø6 mm	0.12 kg
HI026	O ₂	90°	Hose connector Ø6 mm	0.16 kg
HI015	O ₂	In line	G1/4" F R	0.11 kg

